

VERIFICATION OF CALIBRATION REPORT

FEB 01 2010

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130187</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Dan Pasquariello</u>	ID# <u>2620</u> Date <u>1-26-10</u> ✓
A Agency <u>Dillingham Police Dept.</u>	Phone # <u>842-5354</u>
Instrument Location <u>Dillingham Police Department</u> ✓	
B Alco S/N <u>78853</u> Target Value <u>.083</u> ✓ High Pressure <u>650 psi</u>	
Alco Test Value Average <u>.082</u> ✓ 1 st Alco	
<u>.081</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u> ✓	
COB 2/3/10	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Feb, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187 ✓

JANUARY 26, 2010 ✓

OPERATOR'S NAME: ✓

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 26620

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 1234567890

DEPT/AGENCY: DL61

CASE/REPORT: 1234567890

TEST TYPE: U ✓

ALCO TARGET VALUE: .083 ✓

ALCO S/N: 78853 ✓

— BREATH ANALYSIS —

.083 ADJUSTED FOR 29.88 in ✓

ALCO TARGET .082 19:42

BLANK TEST .000 19:43

INTERNAL STANDARD VERIFIED 19:43

ALCO TV 29.88 in .082 19:43

BLANK TEST .000 19:44

SUBJECT SAMPLE .000 19:45

BLANK TEST .000 19:46

ALCO TV 29.88 in .081 19:46

BLANK TEST .000 19:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187 ✓

JANUARY 26, 2010 ✓

TIME 19:40

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 29.88 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~■